

SIA "Bio Re"  
Reģ. Nr. LV50103704121, Artilērijas iela 40G, Rīga, LV-1009, info@biore.lv, tel.:  
29404908

### Price Survey

"The purchase of the continuously working gas analyses equipment"

Riga 09.12.2016.

Nr. CAI-LIFE-2016-12-2

Companies are requested to provide the supply for gas analyses equipment.

specification:

No.	Requested specification	Offered Specification
1	The equipment realize the following gas measurements: CH4 0.5 - 100% vol, CO2 0.5 - 100% vol, H2S 100 - 1000ppm, O2 0.2 - 50% vol, NH3 0 - 5000ppm, H2 0.1 - 30% vol	
2	Equipment for working in continuous operation	
3	The possibility to transfer data to the pilot biogas equipment PLC	
4	Supply voltage 220 VAC 50Hz	
5	The equipment use HID and TCD gas detectors	
6	The machine is equipped with a system for automatic gas sample intake from two points	

The equipment must be delivered until 15.02.2017.

Prepayment will be provided.

The offer proposal has to be delivered personally in Artilerijas street 40G, Riga, LV-1009, or by e-mail: info@biore.lv. Contact person for the survey details: Aigars Atvars, +371 25986724. Price survey specification is available on: [www.biore.lv/iepirkumi.php](http://www.biore.lv/iepirkumi.php).

The offer proposal must be submitted until 6th January 2017.

The price survey is carried out within the project LIFE GRASSSERVICE (**LV LIFE12 BIO/LV/001130**) needs.